

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 1 | <p>Blood Bag Tube Sealer</p> <p>Purpose :To Seal the Tubes of Blood Bag after collection of Blood Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. -Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. *****</p> <ul style="list-style-type: none"> * Notched Seal for easy separation * Fixed sealing head * Designed for heavy duty sealing * RAPID SEAL II seals in 0.5 to 1.2 seconds * "Intelligent" RF Board , accepts and adjusts to all size blood tubing * Retractable splash guard * Easy to clean head protector * Sealing Head Configuration: Fixed Sealing head <p>Rating/Capacity of Eqpt: 220-240VAC Type : Desktop Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab) 1</p> | Nos | 1 | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 2 | <p>Blood Bank Refrigerator</p> <p>Purpose :For Storage of Blood Bags Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. *****</p> <p>* Auto Defrost * Forced - Air Circulation * Temperature control * Safety Control System * Transparency Design * Temperature range (°): 4±1 * Capacity: 158 L * Accuracy: 0.1C * LED Display * Shelf/Inner doors: 4/4</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Floor TOP Accessories: To be supplied with complete set of standard accessories Atleast 6 Shelf Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | No | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 3 | <p>Blood Collection Monitor</p> <p>Purpose :For mixing of blood during collection Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. *****</p> <p>* 8 hours of independent work, * More than 50 donations autonomy * 3-dimensional mixing during blood donation * TFT touch screen 480x272 dots, 16-BIT 65,536 colors * Maximum and minimum flow detection * Visual indication of error / warning / donation status * Small size, * lightweight Designed for mobile use Increased efficiency * Changeable battery system External or internal charging possible * Easy to clean and maintain * Technical Data Power supply 19V DC Battery supply 14,4 V Li-Ion * Battery capacity 3100/3400 mAh or higher * Battery charging time 4 h</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Portable Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | Nos | 3 | | |
| | 3 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 4 | <p>Blood Donation Couch Purpose :For Collection of Blood Main_Feature: * Provides safety and comfort to the donor and the phlebotomist. * Helps reduce the chances of the vasovagal attack by reducing blood pooling in the lower limbs. If a vasovagal attack does occur, the donor's head can be quickly lowered with a rocker switch. * Initiation of motor movement: Via a rocker switch * Upholstery: Soft moulded upholstery, 6 cm thick * Maneuverability: Provided with lockable caster wheels * Seat length: 100 cm * Backrest length: 90 cm * Width: 67 cm * Armrest length: 60 cm * Armrest width: 15 cm * Legrest maximum height: 81 cm * Seat height: 52 cm * Vasovagal attack position height: 67 cm * Backrest maximum height: 111 cm</p> <p>Rating/Capacity of Eqpt: Heavy Duty Type : Floor Top Accessories: To be supplied with complete set of standard accessories including, BCM tray, BP apparatus tray, tray for accessories, telescopic IV stand Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, Korea, Taiwan, China, Pakistan (if quality meets the requirement) Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab) 3</p> | No | 3 | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 5 | <p>Cholecystectomy Laproscopic Hand Instruments Set</p> <p>Purpose :to perform laproscopic surgeries Main_Feature: * Marryland dissecting & grasping forcep rotating with connector pin size 5mm length 36cm(complete), consisting of plastic handel without ratchet.</p> <p>* Bullet Tip dissecting & grassping forcep rotating with connector pin with unipolar coagulation size 5mm length 36cm consisting of Plastic handel without ratchet.</p> <p>* Tooth grassping forcep, rotating with connector pin for unipolar coagulation, size 5mm & 36cm, 2x4 teeth, double action jaw, plastic handel with hemostst style ratchet.</p> <p>* Claw grassping forceps, rotating with connector pin for unipolar coagulation, size 5mm & 36cm ,2x3 teeth, single action jaws, plastic handel with hemostat style ratchet.</p> <p>* Metzenbaum scissors, rotating with connector pin for uni polar coagulation size 5mm length 36cm, blades curved, double action jaws, length of blades 12mm, insulated handel, without ratchet.</p> <p>* Yuhan / Bowl grassping forceps, rotating with connector pin for unipolar coagulation, size 5mm & 36cm , Atraumatic, fenestrated single acion jaws, plastic handel without ratchet.</p> <p>* Alligator grasping forcep rotating with connector pin with unipolar coagulation size 5mm length 36cm alligator jaws double action, Plastic handel without ratchet.</p> <p>* Coagulating and dissecting electrode, L-shaped with connector pin for unipolar coagulation size 5mm, working length 36cm.</p> <p>* Forward-Oblique Telescope 30°, enlarged view, diameter 10 mm, length 31 cm, autoclavable, fiber optic light transmission incorporated, color code: red</p> <p>* 2x VERESS Pneumoperitoneum Needle, with spring-loaded blunt inner cannula, LUER-Lock, autoclavable, diameter 2.1 mm, length 13 cm</p> <p>* Plastic Container, for sterilization and storage, with removable instrument rack and insert tray for storage of up to 12 instruments with diameter 2.5 to 10 mm and up to 6 trocars, perforated, with transparent lid, external dimensions (w x d x h): 532 x 254 x 165 mm</p> <p>* Wire Tray, for the cleaning, sterilization and storage of two rigid endoscopes and one light cable, including holder for light post adaptors, silicone telescope holders and lid, external dimensions (w x d x h) : 487 x 125 x 54 mm, for rigid endoscopes up to diameter 10 mm and working length 32 cm</p> <p>* 2x Trocar, size 6 mm, Including Trocar with pyramidal tip, Cannula, Multifunctional Valve</p> <p>* Speculum for Gall Bladder Extraction, length 6 cm</p> <p>* KOH Macro Needle Holder, ergonomic axial handle with disengageable ratchet, ratchet release on the right side,left curved jaws, with tungsten carbide insert ø 5 mm, length 33 cm</p> <p>* KELLY Grasping Forceps, CLERMONT-FERRAND model, rotating, dismantling, with connector pin for bipolar coagulation, especially suitable for dissection, double action jaws, size 5 mm, length 36 cm, color code: light blue, Consisting of: Plastic Handle, without ratchet, Metal Outer Sheath Forceps Insert</p> <p>* Bipolar High Frequency Cord, length 300 cm,</p> <p>* Fascial Closure Instrument, for subcutaneous ligature of trocar incisions, size 2.8 mm, length 17 cm</p> <p>* Scissors, rotating, with connector pin for unipolar coagulation, size 5 mm, length 36 cm, curved, serrated, spoon blades, length of blades 17 mm, double action jaws, Consisting of: Plastic Handle, without ratchet, with larger contact area Outer Tube, insulated Scissors Insert</p> <p>* OEM: M/s Karl Storz (Preferred)</p> <p>Rating/Capacity of Eqpt: * Heavy Duty Stainless Steel</p> <p>* Autoclaveable</p> <p>* Dull Finish</p> <p>Type : Portable</p> <p>Accessories:</p> <p>Nil</p> <p>Desired Measurements : Nil</p> | Set | 1 | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| | Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : Germany, USA, Europe, UK Warranty : 2 Years Likely Delivery Date : Project: FFH SOMH Karachi(Main OT) | | 1 | | |
| 6 | Dialyzer Reprocessor Purpose :To sterilize, clean, test and affuse reusable dialyzer used in hemodialysis treatment. Main_Feature: * CE Certificate should be attached. * FDA Certificate should be attached. * Manufacturer Authorization letter letter should be attached. * Country of origin and manufacture should be usa/europe/japan only * Sub Distributorship is not acceptable * Refrences of Previous installation should be attached. * Eqpt should be newly manufactured and from current production of year * Service manual with Circuit diagram is must without which installation report and payment will not be released. * All required documents need to be attached failing which will automatically declare ineligible for tender * Demonstration of eqpt is must to the end user. ***** * Fully automated / computerized dialyzer reprocessing system. * High standardization in cleaning, volume measuring, leak testing and chemical disinfecting. * No external dilution / minimize chemical contact. * Separated mixing and volume tanks to minimize cross contamination. * Automatically shutdown after system disinfecting. * Simultaneously and independently reprocess two or more dialyzers. * REPROCESSING PROCESS : Automatic cleaning, volume measuring, leak testing and chemical filling. * Should be able to process all types and brands of dialyzers. * Volume measuring range = 25-300 ml * Volume measuring accuracy + 5%. * Leak test method, low limit setting should be specified. * Input pressure 25-30 psi. * Flow rate 1.5 - 6.0 liters/minute * Water consumption 27 liters/dialyzer * Quantity consumed to be specified for dialyzers. * Loose connector * Dialyzer volume priming failure. * Self test and disinfection interlock. * Priming volume lower than limit. * Incoming water pressure failure. * LCD, backlight with auto shut off. * Data display, reprocessing data. Rating/Capacity of Eqpt: 220-240VAC Type : Desktop Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Construction of Dialysis Room Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date : Project: FFH SOMH Karachi(Dialysis Unit) | Nos | 1 | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 7 | <p>Gel Card System</p> <p>Purpose :For Cross matching in Blood Bank Main_Feature: * CE Certificate should be attached. * FDA Certificate should be attached. * Manufacturer Authorization letter letter should be attached. * Country of origin and manufacture should be usa/europe/japan only * Sub Distributorship is not acceptable * Refrences of Previous installation should be attached. * Eqpt should be newly manufactured and from current production of year * Service manual with Circuit diagram is must without which installation report and payment will not be released. * All required documents need to be attached failing which will automatically declare ineligible for tender * Demonstration of eqpt is must to the end user. *****</p> <p>* Digital Incubator for Processing of Gel Cards - 2 independent incubation zones with capacity of 12 Gel cards and 12 test tubes in each zone. - Digital user interface screen for control and visualization of remaining incubation time and real temperature for reach zone. - Fixed pre Set temp of 37°C - Pre set incubation time of 15 minutes, which can be adjusted by the user. - Audible programmable alarm to notify end of incubation period. - Visual and audible indication of any error or power failure. - Removable trash bin - Quick procedure guide. - Recording of log file - Capacity: 24 Gel cards and 24 testing tube</p> <p>* Digital Centrifuge for Manual Processing of Gel Cards - Simultaneous centrifuge of up to 24 Gel cards - Removable and exchangeable spinning head. - Pre defined centrifugal parameters for gel card processing : acceleration, speed, braking and time - Digital user interface screen for control and visualization of centrifuge speed and time remaining for processing. - Audible programmable alarm to notify end of centrifugal time period - Imbalance system detects incorrect placement of GLE cards on the spinning head - Speed: 990rpm - Time: 9 min</p> <p>* Reader exclusively reads and interprets Gel Cards. - Identification of Gel Cards by Bar code - Reading of agglutination reactions in Gel cards by photo image - Digitalization and processing of card image using system software. - Results interpretation. - Data recording and storage to ensure traceability - Reading time: 2s (cards)</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Desktop Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | Nos | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 8 | <p>Long Sucker Tips Long term Storage at -50C to -86 C</p> <p>Purpose :For establishment of Blood bank Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user.</p> <p>*****</p> <ul style="list-style-type: none"> * Upright Cabinet Type * Cooling Type Direct cooling * Manual Defrost Mode * Noise Level (dB) 50 * Capacity(L/Cu.Ft) 388/13.7 * Temperature Range(°C) -40~-86 * Microprocessor controller * LED display <p>Rating/Capacity of Eqpt: 220-240VAC Type : Floor Top Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | Set | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 9 | <p>Plasma Extractor (Plasma Expressor)</p> <p>Purpose :To separate blood into red cells and plasma Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. ***** * Simple and quick operation * A semi-automatic press for plasma separation from centrifuged whole blood * Equipped with optical sensor, clamping unit and control electronics * Automatically clamps tubing when the first red cells appear in tubing at the end of plasma transfer. One operator can handle about 10 to 20 units. * RBC sensing sensitivity is user-adjustable * Gives audible and visual alarms * Good reproducibility and easy to use. * Up to five units of separators can be powered by one power adapter with use of jumper cables.</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Desk Top Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: NA Expendables/Starter Kits : Items required for installation of eqpt Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | No | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 10 | <p>Plasma Pheresis Machine (Plasma Blaster Freezer)</p> <p>Purpose :To be use in Blood Banking Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user.</p> <p>***** * Upright Cabinet Type * Cooling Type Direct cooling * Manual Defrost Mode * Noise Lever (dB) 50 * Capacity(L/Cu.Ft) 388/13.7 * Temperature Range(°C) -40~-86 * Microprocessor controller * LED display</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Floor top Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | No | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 11 | <p>Platlet Incubator</p> <p>Purpose :For Processing of Blood Bags Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. -Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. *****</p> <p>ζ Provides the controlled temperature environment required for the storage of platelet products. ζ Set Point Temperature Range: +20°C to +35°C ζ Set Point (default) +22°C ζ Electrical Power 100-240V 50/60Hz ζ Maximum Power Draw 352 W ζ Facility Supply Rating 15 A (minimum) branch circuit with earth ground (USA) ζ Power Plug/Power Cord Length NEMA 5-15 hospital-grade 115V 60Hz, 8 to 10 ft (2.4-3.0m) ζ Ambient Operating Temperature +15°C to +35°C +59°F to +95°F ζ Cooling System: AdvanceCore solid-state thermoelectric heat pump technology with dual peltier forced-air circulation ζ Power Switch On/Off ζ Microprocessor with touchscreen input & display ζ Lockable door, password protected settings, optional PIN entry via keypad ζ Controller Sensor RTD, stainless steel ζ Communication Ports USB, Remote alarm contacts, RJ45 (Ethernet), API available for Ethernet connectivity development. ζ Power Failure Alarm ζ High / Low Alarms ζ Tests fully adjustable ζ Door Ajar Alarm fully adjustable ζ Battery Back-up 18 hr ζ Exterior / Interior Finish Bacteria-resistant powder coating</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Floor Top Accessories: To be supplied with Complete Set of Standard Accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Stabilizer Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | Nos | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 12 | <p>Refrigerated Cryofuge</p> <p>Purpose :To be use in Blood Bank Main_Feature: - CE Certificate should be attached. - FDA Certificate should be attached. - Manufacturer Authorization letter letter should be attached. - Sub Distributorship is not acceptable - Refrences of Previous installation should be attached. - Eqpt should be newly manufactured and from current production of year - Service manual with Circuit diagram is must without which installation report and payment will not be released. - All required documents need to be attached failing which will automatically declare ineligible for tender - Demonstration of eqpt is must to the end user. *****</p> <p>* Enhanced ergonomics with push-button convenience * Simplified operation and space-saving design * Quick run set-up and standardization * Temperature set range - 20 to 40°C, adjustable to 1°C * Capacity: 8 x 550 mL blood bags * Maximum speed 5,000 rpm * Maximum RCF 7,295 x g * Run time 99 hours 59 min 59 sec (1 second increment) * Multilingual selection, on-board training videos, user logging, user lock-out, automatic door opening, automatic rotor ID, on screen-display for imbalance, over temperature, stainless steel chamber, guidance display for error messages</p> <p>Rating/Capacity of Eqpt: 220-240VAC Type : Desktop Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK Warranty : 2 Years Likely Delivery Date :</p> <p>Project: FFH SOMH Karachi(Blood Bank/Path Lab)</p> | Nos | 1 | | |
| | 1 | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| 13 | <p>Surgical Instrument Set</p> <p>Purpose :To carry out surgical proceduress in new Gynae OT Main_Feature: *C-Section Set (Qty 2) each set having following Items:</p> <ul style="list-style-type: none"> -Sponge Holder, 2 Pcs -BP Knife Handle No. 3, 1 Pcs -BP Knife Handle No. 4, 1 Pcs -Artery Forceps Curved 6.5", 6 Pcs -Artery Forceps Straight 8", 4 Pcs -Allis Tissue Forceps 8", 2 Pcs -Allis Tissue Forceps 6.5", 2 Pcs -Green Armytage, 4 Pcs -Vulsillum Forceps, 2 Pcs -Dissecting Forceps Tooth 6.5", 1 Pcs -Dissecting Forceps Tooth 8", 1 Pcs -Dissecting Forceps Plain 6.5", 1 Pcs -Dissecting Forceps Plain 8", 1 Pcs -Diathermy Forceps, 1 Pcs -Mayo Scissor Curved 6.5", 1 Pcs -Mayo Scissor Curved 8", 1 Pcs -Mayo Scissors Straight 8", 1 Pcs -Needle Holder 6.5", 1 Pcs -Needle Holder 8", 1 Pcs -Doyen Retractor, 1 Pcs -Outlet Forceps, 1 Set -Right Angle Retractor Single Ended, 1 Set -Kidney Tray 12", 1 Pcs -Kidney Tray 8", 1 Pcs -S.S Bowl 4", 2 Pcs -Towel Clip 5.5", 6 Pcs -Babcock 6", 2 Pcs -Babcock 8", 2 Pcs -Instrument Pin, 1 Pcs -Probe, 1 Pcs -Instrument Tray 12" x 14", 1 Pcs <p>*Laparotomy Set (Qty 1) each set having following Items:</p> <ul style="list-style-type: none"> -BP Handle # 3, # 4, # 7 (Qty3) -Heavy Tooth Forceps (Qty1) -Disecting Tooth Long Forceps (Qty1) -Disecting Plain Forceps (Qty1) -Tooth forceps 7" (Qty1) -Disecting Plain Forceps 6" (Qty1) -Gillies Tooth Forceps 6" (Qty1) -Metz Scissor 7" St (Qty1) -Metz Scissor 8" St (Qty1) -Metz Scissor 8" cvd (Qty2) -Metz Scissor 6" and 9" (Qty2) -Crile Artery Forceps cvd (Qty6) -Crile Artery Forceps St (Qty6) -Mosquito Artery Forceps cvd (Qty6) -Mosquito Artery Forceps St (Qty6) -Debakey Forceps Long (Qty1) -Debakey Forceps 6" (Qty1) -Allis Tissue Forceps Small (Qty4) -Allis Tissue Forceps Long (Qty2) -Moniyan Forceps (Qty2) -Needle holder 6" and 7" (Qty2) -Needle holder 8" and 9" (Qty2) -Spencer vell St (Qty6) -Kocker St (Qty2) -Kocker Cvd (Qty2) -Robbet Artery Forceps Cvd (Qty4) -Sponge Holder (Qty2) -Babcock Large (Qty2) -Yaunkar Suction (Qty1) -Pool Suction (Qty1) | No | 5 | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|---|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| | <ul style="list-style-type: none"> -Legan Back Ret Large (Qty2) -Legan Back Ret Medium (Qty2) -Morris Retractor (Qty2) -Curette double Ended (Qty1) -Towel Clip (Qty6) -Deaver Retractor (Qty4) -Galli Pot 4" (Qty2) -Kidney Dish Large (Qty2) <p>*Hystrectomy Set (Qty 1) each set having following Items:</p> <ul style="list-style-type: none"> -B.P. Handle #3, #4 (Qty2) -Disecting Forcep Plain (Qty1) -Disecting Forcep Tooth (Qty1) -Long Tongue Forcep Plain (Qty1) -Long Tongue Forcep Tooth (Qty1) -Debekey Forcep Long (Qty1) -Debekey Forcep Small (Qty1) -Mayo Scissor 6" St (Qty1) -Mayo Scissor 5" cvd (Qty1) -Mayo Scissor 7" cvd (Qty1) -Needle Holder 5" (Qty1) -Needle holder 6" (Qty1) -Needle holder 7" (Qty1) -Crile A/F cvd (Qty6) -Crile A/F St (Qty4) -Robbert Fine A/F (Qty3) -Robbert A/F (Qty3) -Kocker Forcep St (Qty4) -Kocker Forcep Cvd (Qty4) -Kocker Double Action Forcep St (Qty2) -Kocker Double Action Forcep Cvd (Qty2) -Vulsillum Forcep (Qty2) -Allis Tissue Forcep (Qty6) -Little Wood Tissue Forcep (Qty2) -Sponge Holder (Qty2) -Langen Back Retractor (Qty2) -Morris Retractor (Qty2) -Deaver Retractor Set (Qty5) -Yunker Suction (Qty1) -Mayomectomy Screw (Qty2) -Towel Clip (Qty5) -Kidney Dish 10" (Qty1) -Galli Pot 4" (Qty2) -Spencer Well A/F St 8" (Qty2) -Mosquito Artery Forcep St. (Qty4) -Mosquito Artery Forcep Cvd (Qty4) -Frezier Suction (Qty1) <p>*D&C Set (Qty 1) each set having following Items:</p> <ul style="list-style-type: none"> -BP Handel # 4 (Qty1) -Mayo Scissor St (Qty1) -Metz Scissors cvd (Qty1) -Dissecting Forceps Plain (Qty1) -Dissecting Forceps Tooth (Qty1) -Cirle A /F (Medium) (Qty2) -Vulsellum forceps (Qty2) -Needle Holder (Qty1) -Sponge holder (Qty1) -Bonney Polyps Forceps (Qty1) -Hegar Cervical Dialators (3-14) (Qty6) -Bladder Sound (Qty2) -Utrine Curette (Qty4) -Utrine Sound (Qty1) -Flushing Currete (Qty3) -Sims Speculum (Qty3) -Duck Speculum (Qty3) -Galli pot (Qty1) | | | | |

| Sr# | Item Description/Specification | A/U | Qty Req | Rate per Unit | |
|-----|--|-----|---------|---------------|------------------------|
| | | | | With out GST | With GST if Applicable |
| | -Kidney Dish Large (Qty1) -Towel Clip (Qty5) Rating/Capacity of Eqpt: Heavy Duty, Dull Finish, Stainless Steel, Autoclaveable Type : Portable Accessories: To be supplied with complete Set of Standard Accessories Desired Measurements : Nil Any Other Detail : Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : Pakistan (Export Quality) Warranty : 2 Years Likely Delivery Date : Project: FFH SOMH Karachi(Gynae OT) 5 | | | | |
| 14 | X-Ray Viewing Box (Rh-View Box) Purpose :For use in Blood Bank Main_Feature: * Built-in temperature indicator for easy, accurate monitoring of viewing area. * Viewing area temperature adjusts easily to compensate for ambient temperature changes, (range 45°C to 50°C). * Rh slide reaches temperature of 37° to 39°C in two minutes, eliminating wasted heat up time. * Soft, fluorescent glare-free light provides uniform illumination of slide contents. * View box can be gently rocked using large knob on support cradle. * Sturdy, easy-to-clean extruded aluminum case is compact, utilizes minimal bench space. * Meets Rh slide test standards set by the AABB. * Unit conveniently includes three-wire cord and plug. * Blood typing. * Rh determinations. * Warming slides for gram-stain. * Tissue typing. * Reading enzyme antibody screens in microplate. Rating/Capacity of Eqpt: 220-240VAc Type : Desktop Accessories: To be supplied with complete set of standard accessories Desired Measurements : Nil Any Other Detail : Nil Ground Support Eqpt : Nil Civil Work: Nil Expendables/Starter Kits : Nil Country of Manufacturing : USA, Europe, Japan, UK, Korea, Taiwan, Turkey Warranty : 2 Years Likely Delivery Date : Project: FFH SOMH Karachi(Blood Bank/Path Lab) 1 | no | 1 | | |